

HYDRAJAWS LIMITED

TESTING DIVISION

Company Registration Number: 2230733 England
1 The Courtyard • Roman Way • Coleshill • Birmingham • B46 1HQ
Telephone 01675 430370 • Fax 01675 465950
Website: www.hydr jaws.co.uk • Email: tester@hydr jaws.co.uk

Certificate of Calibration

GAUGE REF. No. HT1174B 18131
0 - 20 kN
GAUGE RANGE 2000/4
MODEL No.

We certify that this Gauge has been inspected and calibrated for accuracy and passed within our limits of plus or minus 2% F.S.D.

Results obtained are as follows:-

MASTER	kN	4.00	8.00	12.00	16.00	20.00
ACTUAL	kN	4.00	8.00	12.00	16.00	20.00

TRACEABILITY

Gauges manufactured to BS EN 837-1

Calibration undertaken using a Budenberg oil operated dead-weight tester type 280H. Serial number 12127 traceability to UKAS via certificate no. 09195. Dated 31.12.10.

Note: In all accordance with BS EN 837-1 this certificate is valid for a period of 12 months from issue.

Accuracy of gauges as stated above cannot be guaranteed should the unit be subjected to mis-use. Gauge will be permanently damaged should maximum load be exceeded.

THIS CERTIFICATE WILL EXPIRE ON 21/02/2012

Customer DPM (London) Ltd

Order No. Jo.Francis Date of Calibration 21/02/2011

Approved Signatory

Date 21/02/2011